

## Nitrile Examination Gloves

**Designed specifically for the dairy market, Glove Plus Milkers nitrile gloves aid mastitis detection as well as help reduce bacterial contamination.**

Manufactured in a dedicated Nitrile factory, this product is 100% Natural Rubber Latex free and is ideal for glove users who have been identified as having allergic reactions to NR Latex proteins (Type 1 reactions).

### Product Key Points

Enhanced grip and feel - **Level 5**  
EN420 Dexterity Test Pass Value



Fully textured palm & fingers

100% NR Latex Free

Extra Protection - Longer Cuff

Rolled cuff prevents 'roll down'



### Product Information

**Type:** Powder free & non-sterile.

**Material:** Nitrile (Synthetic Latex) - Carbocylated Acrylonitrile Butadiene.

**Protein Content:** 100% NR Latex free.

**Powder Content:** Below 2mg/g glove.

**Features:** Ambidextrous, Beaded Cuff, Textured palm & fingers.

**Colour:** Matt Black.

**Packing:** 50 gloves per dispenser, 10 dispensers per carton.

### Dimensions

Glove Size	Product Code	Palm Width (mm)	Length (mm)
Extra Small	MN71300	< 80	295
Small	MN72300	± 80	295
Medium	MN73300	± 95	295
Large	MN74300	± 110	295
Extra Large	MN75300	± 115	295
XX Large	MN76300	± 120	295

### Thickness

Location of Thickness Measurements	Single Wall (mm)
Finger (at 15mm at the tip)	0.15 ± 0.03
Palm (at centre of palm)	0.13 ± 0.03

### Physical Properties

Parameters	Before Aging	After Aging
Force at Break (N)	10.5	8.0
Watertight AQL	0.65	N/A

### Pre-Shipment Quality Inspection

Criteria	Insp Level	AQL
Dimensions	S-2	4.0
Physical Properties	S-2	4.0
1000ml Water Leak Test	G-1	0.65
Visual Inspection	S-4	4.0

### Quality Assurance:

- This product is manufactured in a plant where the quality management system has been independently assessed as conforming to ISO 9001:2008 and ISO 13485:2008/AC2009.



### Product Conformance:

- EU MDD EN455 Parts 1, 2, 3 & 4.
- Manufactured in compliance to EU PPE 89/686/EEC "Protection Against Chemical Hazard"
- EN374 Parts 1, 2 & 3.
- Passed an independent test to ASTM F1671-07 "Resistance of Materials used in Protective Clothing to Penetration by Blood-Borne Pathogens" (viral test).
- EN420 Dexterity Test - Passed to the maximum Level 5.

[www.milkers-gloves.com](http://www.milkers-gloves.com)